
Diamond

Semi Automated Coagulation Analyser



Diamond

Semi Automated Coagulation Analyser

Technical Specifications

Electronic Design and Instrument Intelligence	Micro-processor based technology with flash-memory
Photometrical Source	Solid state, multi-wavelength Optoelectronic device
Measuring Principle	Photometric with clotting algorithms or using absorbance procedures (end-point or kinetic methods, initial rate) for chromogenic and immunological assays. Only clotting algorithms are currently applied in Diamond
QC Program	On board for all parameters, QC Normal and QC Abnormal data storage, editing and reporting with mean, SD and CV calculation.
Patient Data Profile	On board
Operator Guidance	Information on bottom of the display to guide the user through all tasks of routine measurement, calibration or QC handling
Curve Storage	On board to store all curve data points for PT% Activity and FIBRINOGEN measurement
Reading Volumes	Minimum 150µL, Maximum 400 µL
Required Measuring Time	Parameter dependent – Minimum 5 sec, Maximum 400 sec
Data Transfer / Printout	Bluetooth connection for computers or peripheral to facilitate printout or host connection! Optional external printer available (price extra)
Dimension and Weight	W-175mm, D-240mm, H-75mm Weight 1.5Kg
Environmental Operating Conditions	Temperature: +15 to +35°C; Humidity: 20% to 80% uncondensed Altitude: up to 2000 meters; Use in internal environment only
Cuvettes	Dedicated disposable plastic cuvettes, packed in bags of 100/1000



Marketed In India By:

Fugen
B I O M E D

Fugen Biomed
11, Dharmabhakti Complex,
Khushalwadi, Jamalpore,
Navsari-396445, Gujarat

+91 909 909 5565
contact@fugenbiomed.com
www.fugenbiomed.com